

POSTER PRESENTATION

Open Access

Towards a more comprehensive assessment of cardiovascular fitness - magnetic resonance augmented cardiopulmonary exercise testing (MR-CPEX)

Bejal Pandya*, Grzegorz T Kowalik, Daniel S Knight, Oliver Tann, Graham Derrick, Vivek Muthurangu

From 16th Annual SCMR Scientific Sessions
San Francisco, CA, USA. 31 January - 3 February 2013

Background

Assessment of exercise intolerance is important in patients with cardiovascular disease. Traditionally, this is achieved by measuring maximum VO₂ during cardiopulmonary exercise testing (CPEX). However, this cannot discriminate between cardiac output and tissue extraction problems. A better approach may be to assess VO₂ and cardiac output simultaneously (which also allows calculation of tissue extraction). Thus, we developed MR augmented CPEX in which real-time PCMR is performed at the same time as respiratory gas analysis. The purpose of this study was to validate this novel technology.

Methods

Ten volunteers underwent MR-CPEX in a 1.5T scanner using a ramped protocol on an MR compatible Ergometer. All volunteers exercised till exhaustion and the total test period was 9 minutes. Expired gases and respiratory flow data were collected with a calibrated MR compatible respiratory analysis system. Using this data continuous VO₂, VCO₂ and V_e were calculated for the whole test period. Aortic flow was measured continuously during the test period using real-time UNFOLD-SENSE spiral PCMR (spatial resolution: 2.5x2.5 mm, temporal resolution: 30 ms, 16000 frames). Flow data was segmented using a semi-automated technique to calculate cardiac output during exercise. Cardiac output and VO₂ were used to calculate arterio-venous oxygen content gradient (tissue

oxygen extraction). All volunteers also underwent traditional bicycle CPEX for comparison.

Results

MR augmented CPEX was successful in all volunteers with 40% of participants reaching their anaerobic threshold. The maximum workload reached during MR and conventional CPEX was strongly correlated. ($r=0.76$). Mean peak VO₂ during MR-CPEX was 19.3 ± 5.1 ml/min/kg and peak VCO₂ was 19.6 ± 5.5 ml/min/kg. There was an excellent correlation between MR-CPEX peak VO₂ and conventional CPEX ($r=0.84$). During MR-CPEX, mean heart rate rose from 76 ± 14 to 151 ± 25 bpm, with no change in stroke volume. This resulted in mean cardiac output increasing from 3.2 ± 0.5 l/min/m² to 6.6 ± 1.2 l/min/m². Mean peak arterio-venous oxygen gradient calculated from the cardiac output and VO₂ during MR-CPEX was 12ml O₂ per 100ml of blood. Representative ventilation, cardiac output and tissue arterio-venous oxygen gradient curves are shown in figures 1 & 2.

Conclusions

This study shows that MR-CPEX is a viable technique that can provide a comprehensive assessment of all the components of exercise physiology. During MR-CPEX only submaximal exercise is possible, both due to movement limitation and lack of stroke volume augmentation in the supine position. Nevertheless, there was as strong correlation between traditional CPEX and MR-CPEX. This implies that MR-CPEX does measure useful parameters that are linked to maximal exercise. We believe that the ability to fully measure the cardiopulmonary and peripheral response to exercise will

Cardiac MR, Great Ormond Street Hospital and Institute of Cardiovascular Sciences, University College London, London, UK




Figure 1 Ventilation (top) and simultaneous VO₂ (green) /VCO₂ (red) curves during MR-CPEX in a volunteer achieving their anaerobic threshold.




Figure 2 Arterio-venous Oxygen content gradient (top) and Cardiac Output (bottom) during MR-CPEX.

allow better assessment of exercise intolerance in many cardiac diseases.

Funding

British Heart Foundation Project Grant. Dr Vivek Muthurangu and Dr Bejal Pandya.

Published: 30 January 2013

doi:10.1186/1532-429X-15-S1-P58

Cite this article as: Pandya *et al.*: Towards a more comprehensive assessment of cardiovascular fitness - magnetic resonance augmented cardiopulmonary exercise testing (MR-CPEX). *Journal of Cardiovascular Magnetic Resonance* 2013 15(Suppl 1):P58.

**Submit your next manuscript to BioMed Central
and take full advantage of:**

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit

